



Attestation of Conformity

No. ICR Polska/M4400141



**Name and address
of Registered Manufacturer:**

Guizhou tekang electronics co. LTD
Building 1,2,3,4, standard workshops of aluminum and
aluminum processing base, No.149, Yunhuan Road, Baiyun
District, Guiyang City. China

Product name:

Non-contact Infrared Thermometer

Product type/model:

T168

Trade mark:

n/a

This Attestation confirms that the product meets the requirements of the following normative documents and within limits of its documents gives presumption of conformity with essential requirements of Directive 93/42/EEC.

Relevant EC Directive:

Medical Device Directive 93/42/EEC

**Conformity assessment
procedure:**

EC Declaration of Conformity (Annex VII of Directive
93/42/EEC)

Classification:

Class I according Rule 1 of Annex IX of Directive 93/42/EEC

Applied normative documents:

EN 60601-1: 2006
EN 60601-1-2: 2015

**Applied Quality
Management System**

n/a

This AoC will remain valid only if Quality Management System Certificate remains valid.
The assessment process has been carried out in accordance with the program PC-P-07-07.
Evaluation has been carried out in accordance with test report made by:

- Shenzhen Labman Testing Technology Co., Ltd.

No. of test report:

LM-202003064R

Issue date:

25.03.2020

Expiration date:

24.03.2025

The mutual obligations and rights of the certification are regulated by the contract No. ICR
Polska/2020-3125.

This Attestation applies to products having the same attributes (parameters), intended use, that have
been evaluated and meet the requirements of the aforementioned standard.



Director: Rafał Kalinowski

Warsaw, 25. 03. 2020.

ICR Polska Co. Ltd.

ul. Plac Przymierza 6, 03-944 Warszawa
www.icrpolska.com, e-mail: icrpolska@icrqa.com

